

Dentsply Sirona Implants TitaniumBase library compatibility chart

This chart describes the content of Dentsply Sirona Implants TitaniumBase libraries. The libraries are used when designing restorations for TitaniumBase. The libraries can be used for 3Shape and Exocad design software.



Implant system	TitaniumBase	Atlantis FLO (ref)	Atlantis IO FLO (ref)	3Shape Scan Body <small>(only available for 3Shape)</small>	Elos Accurate IO Scan Body	Elos Accurate Analog for Printed Models
Ankylos®	Ankylos C/, C/ high, /X, /X high	L-01 (34583)	IO-L-01* (35235)	DYR Ø 3.5 DYY Ø 4.5 DYB Ø 5.5 DYG Ø 7.0	IO 8B-A SA	PMA-DAN35/70-1
Astra Tech Implant System® EV	EV 3.0 Green - 1mm/2mm	P-01 (34593)	IO-P-01 (35243)	DEG	IO 3B-A SA	PMA-ATE30-1**
	EV 3.6 Purple - 1mm/2mm	P-02 (34594)	IO-P-02 (35244)	DEP	IO 3B-B SA	PMA-ATE36-1**
	EV 4.2 Yellow - 1mm/2mm	P-03 (34595)	IO-P-03 (35245)	DEY	IO 3B-C SA	PMA-ATE42-1**
	EV 4.8 Blue - 1mm/2mm	P-04 (34596)	IO-P-04 (35246)	DEB	IO 3B-D SA	PMA-ATE48-1**
	EV 5.4 Brown - 1mm/2mm	P-05 (34597)	IO-P-05 (35247)	DE5	IO 3B-E SA	PMA-ATE54-1**
Xive® S	Xive D 3.8	J-03 (34577)	IO-J-03 (35232)	DXY	IO 8A-C	PMA-DXI38-1
	Xive D 4.5	J-03 (34577)	IO-J-04 (35264)	DXB	IO 8A-D	PMA-DXI45-1
	Xive D 5.5	J-03 (34577)	IO-J-05 (35265)	DXR	IO 8A-D	PMA-DXI55-1

* The Atlantis IO FLO for Ankylos C/X is for single use only.

** Astra Tech Implant System Lab Abutment Screw EV is designed for use with Astra Tech Implant System Implant Replica EV only and will not fit into the Elos Accurate Analog for Printed Models. Order an extra Abutment Screw EV or use Elos lab screw.

All trademarks and company names are the property of their respective owners.